



Medica Corporation
 5 Oak Park Drive
 Bedford, Massachusetts 01730
 Tel 781 275 4892
 Fax 781 275 2731
 www.medicacorp.com

Products For Health Care

DISTRIBUTOR NAME: FAMECO GmbH

FAMECO CLAIM REF # 211

CONTACT PERSON: Tobias Pfeifer

EasyLyte

DATE: 2- Oktober-2019

WARRANTY FAILURE REPORT

Product Description	Catalog Number	Lot Number	Serial Number	Analyzer S/N	Install Date (MM/DD/YYYY)	Failure Date (MM/DD/YYYY)	Qty	Failure Description
CA Reference	2152	18342		50262CXPA	25- Februar-2019	14- August-2019	1	K slope too high

Comments Shipment to our headquarter address:	FAMECO GmbH Neumarkter Strasse 34, Haus 2 81673 Muenchen Germany
--	---

COMMENTS:

SEP-18-19; 10:12

KALIBRATION

KALIBRIERWERTE

Na 52.97 K 62.97

Cl 51.61 Ca 24.55

SEP-18-19; 10:26

Na/K/Cl/Ca

CDSE
Release per DCN# 6101

Approved/Date
JW/ NBD 4/26/12



EasyLyte®

REF 2124

Na/K/Cl/Ca/Li

ANALYZER

800 mL

Solutions Pack


Kit de solutions

Confezione di soluzioni


Lösungspackung

Embalagem de Soluções


Kit de soluciones




Store at 25°C
4°C



mediacorp.com/support



2018-10-02




18275/1



2020-09-21



Medica Corporation, 5 Oak Park Drive
Bedford, MA 01730-1413 USA




Emergo Europe, Prinsessegracht 20
2514 AP The Hague, The Netherlands

010621-016 R0

MEDICA



(01) 00386-180095-128 GTIN
(10) 18275/1
(11) 181002
(17) 200921
(422) 840
LOT
PROD DATE
EXPIRY
ORIGIN US



Approved/Date
JW/NBD 4/26/12



H315, H319,
P302+P352,

(KORRELATWERTE)

SEP-18-19; 12:08

GERAETEFUNKTN

KALIBRATION

KALIBRIERWERTE

Na 57.02 K 59.25

Cl 47.86 Ca 19.94

Ca NDRG

TAEGL REINIGER
SEP-18-19; 12:15

STANDBY
SEP-18-19; 12:15

KALIBRATION

KALIBRIERWERTE

Na 56.56 K 59.98

Cl 47.76 Ca 20.52

PROBE 034

Na 140.6 K 3.78

Cl 102.0 Ca 1.20

BLUT Na, K, Cl, Ca mmol/l

(KORRELATWERTE)

1 Normal

LOT 19079

Na+ K+ Cl- Li+

EasyLyte		Na+	K+	Cl-	Li+	Ca+
Na/K, Na/K/Cl, Na/K/Li, Na/K/Cl/Li	135-145	3.5-4.1	96-106	0.91-1.07		
Na/K/Ca/pH	134-144	3.5-4.1				1.07-1
Na/K/Cl/Ca/Li (Expand)	135-145	3.5-4.1	97-107	0.90-1.06	0.98-1	
EasyElectrolytes, Rapidchem 744/754	140	3.80	103			
Na/K/Cl, Na/K/Li	138-148	3.4-4.0	99-109	0.85-1.01		
ISE Module 4-channel *	138-148	3.5-4.1	99-109	0.91-1.07		
EasyRA	134-144	3.4-4.0	97-107	0.84-1.00		

TAEGL REINIGER
SEP-11-19; 06:44

STANDBY

SEP-11-19; 06:44

KALIBRATION

KALIBRIERWERTE

Na 50.73 K 65.19

Cl 54.61 Ca 32.86

K HOCH

SEP-11-19; 06:48

TAEGL REINIGER
SEP-11-19; 06:51

STANDBY

SEP-11-19; 06:51

KALIBRATION

KALIBRIERWERTE

Na 53.21 K 63.05

Cl 51.93 Ca 28.84

SEP-11-19; 06:58

ANALYSE
NAME PATIENT

PROBE 001

MEDICA


EasyLyte®

REF 2101

K+ Electrode
Électrode de potassium
Elettrodo Potassio
Kalium Elektrode
Electrodo de K+
Eléctrodo K+
K+ Elektrodu

LOT 19221

IN 2020-08-09
install by

 2019-08-09



(01) 00386100002961
(10) 19221
(11) 190809
(15) 200809
(422) 840

